

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00430 Calib. Date: 08/26/2011 Calib. No.: 00018  
Certification File No.: 00431 Cert. Date: 08/26/2011 Cert. No.: 00012  
Linearity File No.: 00432 Lin. Date: 08/26/2011 Lin. No.: 00012  
Solution File No.: 00440 Soln. Date: 10/20/2011 Soln. No.: 00106  
Sequential File No.: 00440 File Date: 10/20/2011

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061  
Control Solution %: 0.100% Expires: 12/16/2012  
Solution Control Lot: 10L084 Bottle No.: 0132

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 09:36D        |                               |                           |
| Control 1 EC      | 0.101%         | 09:36D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.101%         | 09:36D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 09:37D        |                               |                           |
| Control 2 EC      | 0.100%         | 09:38D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.101%         | 09:38D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 09:38D        |                               |                           |
| Control 3 EC      | 0.101%         | 09:39D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.101%         | 09:39D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 09:39D        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HOPSTEN

First Name: GARY

MI: W

Signature:  #351

Badge No.: 351

Date: 10/20/2011