

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00785	Calib. Date: 08/20/2015
Certification File No.:	00786	Calib. No.: 00031
Linearity File No.:	00787	Cert. Date: 08/20/2015
Solution File No.:	00805	Cert. No.: 00021
Sequential File No.:	00805	Lin. Date: 08/20/2015
		Lin. No.: 00021
		Soln. Date: 10/20/2015
		Soln. No.: 00171
		File Date: 10/20/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0061
Solution Control Lot:	14H132	Expires: 08/22/2016
		Bottle No.: 0026

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:24D		
Control 1 EC	0.100%	07:24D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.102%	07:24D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:25D		
Control 2 EC	0.100%	07:26D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:26D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:26D		
Control 3 EC	0.101%	07:27D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.101%	07:27D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:28D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Badge No.: 421

Date: 10/20/2015