

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00858 Calib. Date: 07/11/2016 Calib. No.: 00033
Certification File No.: 00859 Cert. Date: 07/11/2016 Cert. No.: 00023
Linearity File No.: 00860 Lin. Date: 07/11/2016 Lin. No.: 00023
Solution File No.: 00878 Soln. Date: 10/20/2016 Soln. No.: 00186
Sequential File No.: 00878 File Date: 10/20/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0060
Control Solution %: 0.100% Expires: 07/06/2018
Solution Control Lot: 16150 Bottle No.: 1346

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:33D		
Control 1 EC	0.100%	11:33D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:34D		
Control 2 EC	0.100%	11:35D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:35D		
Control 3 EC	0.100%	11:36D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:37D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 10/20/2016