

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00327	Calib. Date: 07/02/2012
Certification File No.:	00328	Calib. No.: 00023
Linearity File No.:	00329	Cert. Date: 07/02/2012
Solution File No.:	00343	Cert. No.: 00015
Sequential File No.:	00343	Lin. Date: 07/02/2012
		Lin. No.: 00015
		Soln. Date: 10/23/2012
		Soln. No.: 00114
		File Date: 10/23/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0079
Solution Control Lot:	11I095	Expires: 09/15/2013
		Bottle No.: 0426

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:38D		
Control 1 EC	0.099%	15:39D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:39D		
Control 2 EC	0.098%	15:40D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:40D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:40D		
Control 3 EC	0.099%	15:41D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:41D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:42D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: \_\_\_\_\_



Badge No.: 351

Date: 10/23/2012

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTL-0014  
 Location: SOUTH BRUNSWICK POLICE  
 Calibration File No.: 00327 Calib. Date: 07/02/2012 Calib. No.: 00023  
 Certification File No.: 00328 Cert. Date: 07/02/2012 Cert. No.: 00015  
 Linearity File No.: 00329 Lin. Date: 07/02/2012 Lin. No.: 00015  
 Solution File No.: 00342 Soln. Date: 10/23/2012 Soln. No.: 00113  
 Sequential File No.: 00342 File Date: 10/23/2012

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0079  
 Control Solution %: 0.100% Expires: 09/15/2013  
 Solution Control Lot: 111095 Bottle No.: 0436

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:00D		
Control 1 EC	-.---%	13:00D	21.8°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	13:00D	21.8°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: \_\_\_\_\_



Badge No.: 351

Date: 10/23/2012