

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00454 Calib. Date: 07/11/2016 Calib. No.: 00034
Certification File No.: 00455 Cert. Date: 07/11/2016 Cert. No.: 00025
Linearity File No.: 00456 Lin. Date: 07/11/2016 Lin. No.: 00025
Solution File No.: 00465 Soln. Date: 10/27/2016 Soln. No.: 00180
Sequential File No.: 00465 File Date: 10/27/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0080
Control Solution %: 0.100% Expires: 07/06/2018
Solution Control Lot: 16150 Bottle No.: 1384

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:48D		
Control 1 EC	0.099%	07:49D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:49D		
Control 2 EC	0.098%	07:50D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:51D		
Control 3 EC	0.099%	07:51D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	07:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:52D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 10/27/2016