

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTL-0014  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00406 Calib. Date: 09/30/2014 Calib. No.: 00029  
Certification File No.: 00407 Cert. Date: 09/30/2014 Cert. No.: 00021  
Linearity File No.: 00408 Lin. Date: 09/30/2014 Lin. No.: 00021  
Solution File No.: 00411 Soln. Date: 11/03/2014 Soln. No.: 00146  
Sequential File No.: 00411 File Date: 11/03/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0079  
Control Solution %: 0.100% Expires: 11/12/2015  
Solution Control Lot: 13K125 Bottle No.: 1327

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:36S		
Control 1 EC	0.101%	10:36S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	10:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:37S		
Control 2 EC	0.100%	10:37S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:38S		
Control 3 EC	0.100%	10:39S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:39S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_



Badge No.: 421

Date: 11/03/2014

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00406	Calib. No.: 00029
Certification File No.:	00407	Cert. No.: 00021
Linearity File No.:	00408	Lin. No.: 00021
Solution File No.:	00410	Soln. No.: 00145
Sequential File No.:	00410	File Date: 11/03/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0079
Solution Control Lot:	13K125	Expires: 11/12/2015
		Bottle No.: 1327

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:30S		
Control 1 EC	0.101%	08:30S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.009%	08:31S		
Control 2 EC	0.097%	08:32S	34.0°C	AMBIENT AIR CHECK ERR
Control 2 IR	-.---%	08:32S	34.0°C	AMBIENT AIR CHECK ERR
Ambient Air Blank	-.---%	08:32S		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_ 

Badge No.: 421

Date: 11/03/2014