

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00705	Calib. No.: 00029
Certification File No.:	00706	Cert. No.: 00019
Linearity File No.:	00707	Lin. No.: 00019
Solution File No.:	00717	Soln. No.: 00156
Sequential File No.:	00717	File Date: 11/03/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0061
Solution Control Lot:	13K125	Expires: 11/12/2015
		Bottle No.: 1326

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:13S		
Control 1 EC	0.099%	08:14S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:14S		
Control 2 EC	0.099%	08:15S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	08:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:16S		
Control 3 EC	0.098%	08:16S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.095%	08:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:17S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 11/03/2014