

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00343	Calib. Date: 06/21/2010
Certification File No.:	00344	Calib. No.: 00013
Linearity File No.:	00345	Cert. Date: 06/21/2010
Solution File No.:	00371	Cert. No.: 00009
Sequential File No.:	00371	Lin. Date: 06/21/2010
		Lin. No.: 00009
		Soln. Date: 11/10/2010
		Soln. No.: 00089
		File Date: 11/10/2010

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUF S3-0061
Control Solution %: 0.100%		Expires: 09/18/2011
Solution Control Lot: 091069		Bottle No.: 0399

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:21S		
Control 1 EC	0.101%	15:22S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	15:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:23S		
Control 2 EC	0.100%	15:23S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	15:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:24S		
Control 3 EC	0.100%	15:25S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature:  #351

Badge No.: 351

Date: 11/10/2010