

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00263 Calib. Date: 06/24/2009 Calib. No.: 00011
Certification File No.: 00264 Cert. Date: 06/24/2009 Cert. No.: 00007
Linearity File No.: 00265 Lin. Date: 06/24/2009 Lin. No.: 00007
Solution File No.: 00291 Soln. Date: 11/11/2009 Soln. No.: 00073
Sequential File No.: 00291 File Date: 11/11/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0080
Control Solution %: 0.100% Expires: 12/01/2010
Solution Control Lot: 08L061 Bottle No.: 1073

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:59S		
Control 1 EC	0.099%	15:59S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:00S		
Control 2 EC	0.099%	16:01S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:01S		
Control 3 EC	0.099%	16:02S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:03S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROGERS

First Name: MICHAEL

MI: J.

Signature: 

Badge No.: 258

Date: 11/11/2009