

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00305 Calib. Date: 08/26/2011 Calib. No.: 00021
Certification File No.: 00306 Cert. Date: 08/26/2011 Cert. No.: 00013
Linearity File No.: 00307 Lin. Date: 08/26/2011 Lin. No.: 00013
Solution File No.: 00311 Soln. Date: 11/15/2011 Soln. No.: 00096
Sequential File No.: 00311 File Date: 11/15/2011

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0079
Control Solution %: 0.100% Expires: 12/16/2012
Solution Control Lot: 10L084 Bottle No.: 0146

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:04S		
Control 1 EC	0.103%	16:05S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	16:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:06S		
Control 2 EC	0.101%	16:06S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	16:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:07S		
Control 3 EC	0.101%	16:07S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	16:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:08S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: 

Badge No.: 351

Date: 11/15/2011