

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00430 Calib. Date: 08/26/2011 Calib. No.: 00018
Certification File No.: 00431 Cert. Date: 08/26/2011 Cert. No.: 00012
Linearity File No.: 00432 Lin. Date: 08/26/2011 Lin. No.: 00012
Solution File No.: 00443 Soln. Date: 11/15/2011 Soln. No.: 00107
Sequential File No.: 00443 File Date: 11/15/2011

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061
Control Solution %: 0.100% Expires: 12/16/2012
Solution Control Lot: 10L084 Bottle No.: 0141

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:45S		
Control 1 EC	0.101%	13:46S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:46S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:46S		
Control 2 EC	0.101%	13:47S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:48S		
Control 3 EC	0.101%	13:48S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	13:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:49S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 11/15/2011