

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00858	Calib. Date: 07/11/2016
Certification File No.:	00859	Calib. No.: 00033
Linearity File No.:	00860	Cert. Date: 07/11/2016
Solution File No.:	00883	Cert. No.: 00023
Sequential File No.:	00883	Lin. Date: 07/11/2016
		Lin. No.: 00023
		Soln. Date: 11/15/2016
		Soln. No.: 00187
		File Date: 11/15/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0060
Solution Control Lot:	16150	Expires: 07/06/2018
		Bottle No.: 1260

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:51S		
Control 1 EC	0.100%	11:51S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:52S		
Control 2 EC	0.100%	11:53S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:53S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:53S		
Control 3 EC	0.099%	11:54S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:55S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 11/15/2016