

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00433	Calib. Date: 08/20/2015
Certification File No.:	00434	Cert. No.: 00032
Linearity File No.:	00435	Cert. Date: 08/20/2015
Solution File No.:	00439	Lin. No.: 00023
Sequential File No.:	00439	Soln. No.: 00164
		File Date: 11/17/2015

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUH S3-0079
Control Solution %:	0.100%		Expires: 08/04/2017
Solution Control Lot:	15H140		Bottle No.: 0096

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:42S		
Control 1 EC	0.100%	14:42S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:42S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:43S		
Control 2 EC	0.099%	14:43S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:43S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:44S		
Control 3 EC	0.099%	14:45S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:45S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:45S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_

Badge No.: 421

Date: 11/17/2015