

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00327 Calib. Date: 07/02/2012 Calib. No.: 00023
Certification File No.: 00328 Cert. Date: 07/02/2012 Cert. No.: 00015
Linearity File No.: 00329 Lin. Date: 07/02/2012 Lin. No.: 00015
Solution File No.: 00344 Soln. Date: 11/20/2012 Soln. No.: 00115
Sequential File No.: 00344 File Date: 11/20/2012

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0079
Control Solution %: 0.100% Expires: 06/28/2014
Solution Control Lot: 12F102 Bottle No.: 0405

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:30S		
Control 1 EC	0.100%	14:30S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:30S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:31S		
Control 2 EC	0.099%	14:32S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:32S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:32S		
Control 3 EC	0.099%	14:33S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: *[Handwritten Signature]*

Badge No.: 351

Date: 11/20/2012