

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00406 Calib. Date: 09/30/2014 Calib. No.: 00029
Certification File No.: 00407 Cert. Date: 09/30/2014 Cert. No.: 00021
Linearity File No.: 00408 Lin. Date: 09/30/2014 Lin. No.: 00021
Solution File No.: 00413 Soln. Date: 12/01/2014 Soln. No.: 00148
Sequential File No.: 00413 File Date: 12/01/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0079
Control Solution %: 0.100% Expires: 11/12/2015
Solution Control Lot: 13K125 Bottle No.: 1329

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:01S		
Control 1 EC	0.103%	12:01S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:02S		
Control 2 EC	0.100%	12:03S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	12:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:03S		
Control 3 EC	0.100%	12:04S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:05S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature:  #921

Badge No.: 421

Date: 12/01/2014

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00406	Calib. Date: 09/30/2014
Certification File No.:	00407	Calib. No.: 00029
Linearity File No.:	00408	Cert. Date: 09/30/2014
Solution File No.:	00412	Cert. No.: 00021
Sequential File No.:	00412	Lin. Date: 09/30/2014
		Lin. No.: 00021
		Soln. Date: 12/01/2014
		Soln. No.: 00147
		File Date: 12/01/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0079
Solution Control Lot:	13K125	Expires: 11/12/2015
		Bottle No.: 1329

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:55S		
Control 1 EC	0.101%	10:55S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	10:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:56S		
Control 2 EC	0.095%	10:57S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	10:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:58S		
Control 3 EC	0.095%	10:58S	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.093%	10:58S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	10:59S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____  _____

Badge No.: 421

Date: 12/01/2014