

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00705 Calib. Date: 09/30/2014 Calib. No.: 00029
Certification File No.: 00706 Cert. Date: 09/30/2014 Cert. No.: 00019
Linearity File No.: 00707 Lin. Date: 09/30/2014 Lin. No.: 00019
Solution File No.: 00723 Soln. Date: 12/01/2014 Soln. No.: 00157
Sequential File No.: 00723 File Date: 12/01/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061
Control Solution %: 0.100% Expires: 11/12/2015
Solution Control Lot: 13K125 Bottle No.: 1328

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:57S		
Control 1 EC	0.099%	07:58S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	07:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:59S		
Control 2 EC	0.100%	07:59S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:00S		
Control 3 EC	0.100%	08:01S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:01S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

PH. [Signature] # 421

Badge No.: 421

Date: 12/01/2014