

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARNK-0074  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00631 Calib. Date: 10/17/2013 Calib. No.: 00027  
Certification File No.: 00632 Cert. Date: 10/17/2013 Cert. No.: 00017  
Linearity File No.: 00633 Lin. Date: 10/17/2013 Lin. No.: 00017  
Solution File No.: 00645 Soln. Date: 12/12/2013 Soln. No.: 00143  
Sequential File No.: 00645 File Date: 12/12/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0060  
Control Solution %: 0.100% Expires: 03/12/2015  
Solution Control Lot: 13C111 Bottle No.: 0134

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:52S		
Control 1 EC	0.100%	18:53S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	18:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:53S		
Control 2 EC	0.100%	18:54S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	18:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:55S		
Control 3 EC	0.101%	18:55S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.101%	18:55S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:56S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: Sgt. Gary A. #351

Badge No.: 351

Date: 12/12/2013