

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00433 Calib. Date: 08/20/2015 Calib. No.: 00032
Certification File No.: 00434 Cert. Date: 08/20/2015 Cert. No.: 00023
Linearity File No.: 00435 Lin. Date: 08/20/2015 Lin. No.: 00023
Solution File No.: 00440 Soln. Date: 12/15/2015 Soln. No.: 00165
Sequential File No.: 00440 File Date: 12/15/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0079
Control Solution %: 0.100% Expires: 08/04/2017
Solution Control Lot: 15H140 Bottle No.: 0086

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:25S		
Control 1 EC	0.099%	12:26S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:26S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:27S		
Control 2 EC	0.099%	12:27S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:27S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:28S		
Control 3 EC	0.099%	12:28S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:28S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 12/15/2015