

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00327	Calib. No.: 00023
Certification File No.:	00328	Cert. No.: 00015
Linearity File No.:	00329	Lin. No.: 00015
Solution File No.:	00345	Soln. No.: 00116
Sequential File No.:	00345	File Date: 12/20/2012

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUH S3-0079
Control Solution %: 0.100%		Expires: 06/28/2014
Solution Control Lot: 12F102		Bottle No.: 0401

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:31S		
Control 1 EC	0.100%	13:32S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:32S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:32S		
Control 2 EC	0.099%	13:33S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:34S		
Control 3 EC	0.099%	13:34S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:34S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:35S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____

[Handwritten Signature] #351

Badge No.: 351

Date: 12/20/2012