

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00272 Calib. Date: 12/03/2010 Calib. No.: 00017
Certification File No.: 00273 Cert. Date: 12/03/2010 Cert. No.: 00011
Linearity File No.: 00274 Lin. Date: 12/03/2010 Lin. No.: 00011
Solution File No.: 00276 Soln. Date: 12/28/2010 Soln. No.: 00082
Sequential File No.: 00276 File Date: 12/28/2010

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0080
Control Solution %: 0.100% Expires: 09/18/2011
Solution Control Lot: 091069 Bottle No.: 0395

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:02S		
Control 1 EC	0.102%	10:02S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:03S		
Control 2 EC	0.100%	10:04S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:04S		
Control 3 EC	0.101%	10:05S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:06S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____

Badge No.: 351

Date: 12/28/2010