

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00406	Calib. Date:	09/30/2014
Certification File No.:	00407	Cert. No.:	00029
Linearity File No.:	00408	Cert. Date:	09/30/2014
Solution File No.:	00414	Lin. Date:	09/30/2014
Sequential File No.:	00414	Soln. Date:	12/30/2014
		File Date:	12/30/2014

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUH S3-0079
Control Solution %:	0.100%			Expires:	08/22/2016
Solution Control Lot:	14H132			Bottle No.:	0050

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:46S		
Control 1 EC	0.102%	13:47S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	13:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:48S		
Control 2 EC	0.100%	13:48S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	13:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:49S		
Control 3 EC	0.099%	13:50S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:51S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: HOLSTEN

First Name: GARY

MI: W

Signature: _____



Badge No.: 351

Date: 12/30/2014