

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00454	Calib. No.: 00034
Certification File No.:	00455	Cert. No.: 00025
Linearity File No.:	00456	Lin. No.: 00025
Solution File No.:	00468	Soln. No.: 00182
Sequential File No.:	00468	File Date: 12/30/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0080
Solution Control Lot:	16150	Expires: 07/06/2018
		Bottle No.: 1469

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:01S		
Control 1 EC	0.100%	11:01S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:02S		
Control 2 EC	0.099%	11:03S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:04S		
Control 3 EC	0.100%	11:04S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:05S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_



Badge No.: 421

Date: 12/30/2016