

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00454	Calib. Date:	07/11/2016
Certification File No.:	00455	Cert. Date:	07/11/2016
Linearity File No.:	00456	Lin. Date:	07/11/2016
Solution File No.:	00466	Soln. Date:	11/23/2016
Sequential File No.:	00466	File Date:	11/23/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUH S3-0080
Solution Control Lot:	16150	Expires:	07/06/2018
		Bottle No.:	1460

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:53S		
Control 1 EC	0.101%	07:54S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:55S		
Control 2 EC	0.099%	07:55S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:56S		
Control 3 EC	0.099%	07:57S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:57S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Badge No.: 421

Date: 11/23/2016