

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00537	Calib. No.: 00039
Certification File No.:	00538	Cert. No.: 00030
Linearity File No.:	00539	Lin. No.: 00030
Solution File No.:	00543	Soln. No.: 00212
Sequential File No.:	00543	File Date: 12/05/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0080
Solution Control Lot:	18150	Expires: 05/10/2020
		Bottle No.: 1294

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:20S		
Control 1 EC	0.100%	08:21S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:22S		
Control 2 EC	0.099%	08:22S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:23S		
Control 3 EC	0.099%	08:24S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:24S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BLAKE

First Name: JESSE

MI: R

Signature:  #421

Badge No.: 421

Date: 12/05/2018