

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTL-0014  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00469 Calib. Date: 01/04/2017 Calib. No.: 00035  
Certification File No.: 00470 Cert. Date: 01/04/2017 Cert. No.: 00026  
Linearity File No.: 00471 Lin. Date: 01/04/2017 Lin. No.: 00026  
Solution File No.: 00473 Soln. Date: 02/01/2017 Soln. No.: 00184  
Sequential File No.: 00473 File Date: 02/01/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0080  
Control Solution %: 0.100% Expires: 07/06/2018  
Solution Control Lot: 16150 Bottle No.: 1383

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:00S		
Control 1 EC	0.100%	08:00S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	08:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:01S		
Control 2 EC	0.098%	08:02S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:02S		
Control 3 EC	0.098%	08:03S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:04S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_



Badge No.: 421

Date: 02/01/2017