

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00893	Calib. No.: 00034
Certification File No.:	00894	Cert. No.: 00024
Linearity File No.:	00895	Lin. No.: 00024
Solution File No.:	00898	Soln. No.: 00190
Sequential File No.:	00898	File Date: 02/01/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0060
Solution Control Lot:	16150	Expires: 07/06/2018
		Bottle No.: 1257

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:59S		
Control 1 EC	0.099%	07:59S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:00S		
Control 2 EC	0.099%	08:01S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:02S		
Control 3 EC	0.099%	08:03S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:03S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:04S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 02/01/2017