

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00940	Calib. Date: 11/24/2017
Certification File No.:	00941	Calib. No.: 00036
Linearity File No.:	00942	Cert. Date: 11/24/2017
Solution File No.:	00951	Cert. No.: 00026
Sequential File No.:	00951	Lin. Date: 11/24/2017
		Lin. No.: 00026
		Soln. Date: 01/23/2018
		Soln. No.: 00205
		File Date: 01/23/2018

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0060
Solution Control Lot:	17110	Expires: 03/20/2019
		Bottle No.: 1111

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:55S		
Control 1 EC	0.096%	07:55S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	07:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:56S		
Control 2 EC	0.098%	07:57S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:58S		
Control 3 EC	0.095%	07:59S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	07:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Badge No.: 421

Date: 01/23/2018