

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00893 Calib. Date: 01/04/2017 Calib. No.: 00034
Certification File No.: 00894 Cert. Date: 01/04/2017 Cert. No.: 00024
Linearity File No.: 00895 Lin. Date: 01/04/2017 Lin. No.: 00024
Solution File No.: 00908 Soln. Date: 04/27/2017 Soln. No.: 00193
Sequential File No.: 00908 File Date: 04/27/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0060
Control Solution %: 0.100% Expires: 07/06/2018
Solution Control Lot: 16150 Bottle No.: 1331

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:18D		
Control 1 EC	0.099%	12:18D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:19D		
Control 2 EC	0.099%	12:20D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:21D		
Control 3 EC	0.100%	12:21D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:22D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 04/27/2017