

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0074
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00912	Calib. Date:	05/31/2017
Certification File No.:	00913	Cert. Date:	05/31/2017
Linearity File No.:	00914	Lin. Date:	05/31/2017
Solution File No.:	00921	Soln. Date:	06/29/2017
Sequential File No.:	00921	File Date:	06/29/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUF S3-0061
Solution Control Lot:	16150	Expires:	07/06/2018
		Bottle No.:	1427

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:04D		
Control 1 EC	0.101%	13:04D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:05D		
Control 2 EC	0.100%	13:06D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:06D		
Control 3 EC	0.100%	13:07D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:08D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_

Badge No.: 421

Date: 06/29/2017