

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00958 Calib. Date: 03/27/2018 Calib. No.: 00039
Certification File No.: 00959 Cert. Date: 03/27/2018 Cert. No.: 00027
Linearity File No.: 00960 Lin. Date: 03/27/2018 Lin. No.: 00027
Solution File No.: 00974 Soln. Date: 07/11/2018 Soln. No.: 00212
Sequential File No.: 00974 File Date: 07/11/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUF S3-0061
Control Solution %: 0.100% Expires: 12/11/2019
Solution Control Lot: 17420 Bottle No.: 0194

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:51D		
Control 1 EC	0.100%	07:52D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:52D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:52D		
Control 2 EC	0.100%	07:53D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:53D		
Control 3 EC	0.100%	07:54D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 07/11/2018