

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00912	Calib. No.: 00035
Certification File No.:	00913	Cert. No.: 00025
Linearity File No.:	00914	Lin. No.: 00025
Solution File No.:	00924	Soln. No.: 00197
Sequential File No.:	00924	File Date: 07/26/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0061
Solution Control Lot:	17110	Expires: 03/20/2019
		Bottle No.: 0790

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:56D		
Control 1 EC	0.101%	07:57D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	07:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:57D		
Control 2 EC	0.102%	07:58D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:59D		
Control 3 EC	0.101%	07:59D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	07:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_

Badge No.: 421

Date: 07/26/2017