

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00980	Calib. Date: 09/05/2018
Certification File No.:	00981	Cert. No.: 00040
Linearity File No.:	00982	Cert. Date: 09/05/2018
Solution File No.:	00985	Lin. Date: 09/05/2018
Sequential File No.:	00985	Lin. No.: 00028
		Soln. Date: 09/26/2018
		Soln. No.: 00215
		File Date: 09/26/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0060
Solution Control Lot:	17420	Expires: 12/11/2019
		Bottle No.: 0083

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:47D		
Control 1 EC	0.101%	08:47D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:48D		
Control 2 EC	0.100%	08:48D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:49D		
Control 3 EC	0.099%	08:50D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	08:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:50D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Badge No.: 421

Date: 09/26/2018