

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00980	Calib. Date: 09/05/2018
Certification File No.:	00981	Cert. No.: 00040
Linearity File No.:	00982	Cert. Date: 09/05/2018
Solution File No.:	00992	Lin. Date: 09/05/2018
Sequential File No.:	00992	Lin. No.: 00028
		Soln. Date: 10/24/2018
		Soln. No.: 00216
		File Date: 10/24/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0060
Solution Control Lot:	17420	Expires: 12/11/2019
		Bottle No.: 0968

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:50D		
Control 1 EC	-.---%	12:51D	21.5°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	12:51D	21.5°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Badge No.: 421

Date: 10/24/2018

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0074
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00980 **Calib. Date:** 09/05/2018 **Calib. No.:** 00040
Certification File No.: 00981 **Cert. Date:** 09/05/2018 **Cert. No.:** 00028
Linearity File No.: 00982 **Lin. Date:** 09/05/2018 **Lin. No.:** 00028
Solution File No.: 00993 **Soln. Date:** 10/24/2018 **Soln. No.:** 00217
Sequential File No.: 00993 **File Date:** 10/24/2018

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUF S3-0060
Control Solution %: 0.100% **Expires:** 12/11/2019
Solution Control Lot: 17420 **Bottle No.:** 0968

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:35D		
Control 1 EC	0.100%	15:36D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:36D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:37D		
Control 2 EC	0.100%	15:37D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:38D		
Control 3 EC	0.100%	15:38D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:38D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:39D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Badge No.: 421

Date: 10/24/2018