

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARNK-0074
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00980	Calib. No.: 00040
Certification File No.:	00981	Cert. No.: 00028
Linearity File No.:	00982	Lin. No.: 00028
Solution File No.:	00999	Soln. No.: 00218
Sequential File No.:	00999	File Date: 11/21/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUF S3-0060
Solution Control Lot:	18150	Expires: 05/10/2020
		Bottle No.: 1117

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:17S		
Control 1 EC	0.099%	20:18S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	20:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:18S		
Control 2 EC	0.099%	20:19S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:19S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:20S		
Control 3 EC	0.098%	20:20S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:20S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:21S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_

Badge No.: 421

Date: 11/21/2018