

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00537	Calib. No.: 00039
Certification File No.:	00538	Cert. No.: 00030
Linearity File No.:	00539	Lin. No.: 00030
Solution File No.:	00545	Soln. No.: 00214
Sequential File No.:	00545	File Date: 01/02/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0080
Solution Control Lot:	18150	Expires: 05/10/2020
		Bottle No.: 1291

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:35S		
Control 1 EC	0.100%	14:36S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:36S		
Control 2 EC	0.098%	14:37S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:38S		
Control 3 EC	0.098%	14:38S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:39S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Badge No.: 421

Date: 01/02/2019

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00537	Calib. Date: 09/05/2018
Certification File No.:	00538	Calib. No.: 00039
Linearity File No.:	00539	Cert. Date: 09/05/2018
Solution File No.:	00544	Cert. No.: 00030
Sequential File No.:	00544	Lin. Date: 09/05/2018
		Lin. No.: 00030
		Soln. Date: 01/02/2019
		Soln. No.: 00213
		File Date: 01/02/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0080
Solution Control Lot:	18150	Expires: 05/10/2020
		Bottle No.: 1291

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:45S		
Control 1 EC	-.---%	12:46S	21.0°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	12:46S	21.0°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Badge No.: 421

Date: 01/02/2019