

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00497 Calib. Date: 11/24/2017 Calib. No.: 00037
Certification File No.: 00498 Cert. Date: 11/24/2017 Cert. No.: 00028
Linearity File No.: 00499 Lin. Date: 11/24/2017 Lin. No.: 00028
Solution File No.: 00505 Soln. Date: 01/19/2018 Soln. No.: 00198
Sequential File No.: 00505 File Date: 01/19/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0080
Control Solution %: 0.100% Expires: 03/20/2019
Solution Control Lot: 17110 Bottle No.: 1070

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:49S		
Control 1 EC	0.101%	15:49S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:50S		
Control 2 EC	0.099%	15:51S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:52S		
Control 3 EC	0.099%	15:52S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:53S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 01/19/2018