

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00497	Calib. Date: 11/24/2017
Certification File No.:	00498	Calib. No.: 00037
Linearity File No.:	00499	Cert. Date: 11/24/2017
Solution File No.:	00509	Cert. No.: 00028
Sequential File No.:	00509	Lin. Date: 11/24/2017
		Lin. No.: 00028
		Soln. Date: 01/30/2018
		Soln. No.: 00199
		File Date: 01/30/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0080
Solution Control Lot:	17110	Expires: 03/20/2019
		Bottle No.: 0921

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:05S		
Control 1 EC	0.100%	15:06S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:06S		
Control 2 EC	0.098%	15:07S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:08S		
Control 3 EC	0.100%	15:08S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:09S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature:  #421

Badge No.: 421

Date: 01/30/2018