

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00524	Calib. Date: 03/27/2018
Certification File No.:	00525	Cert. No.: 00038
Linearity File No.:	00526	Cert. Date: 03/27/2018
Solution File No.:	00528	Lin. No.: 00029
Sequential File No.:	00528	Lin. Date: 03/27/2018
		Soln. No.: 00202
		Soln. Date: 04/25/2018
		File Date: 04/25/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0079
Solution Control Lot:	17110	Expires: 03/20/2019
		Bottle No.: 0951

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:15D		
Control 1 EC	0.100%	08:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:16D		
Control 2 EC	0.099%	08:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:17D		
Control 3 EC	0.098%	08:18D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:18D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Badge No.: 421

Date: 04/25/2018