

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00469	Calib. Date: 01/04/2017
Certification File No.:	00470	Calib. No.: 00035
Linearity File No.:	00471	Cert. Date: 01/04/2017
Solution File No.:	00481	Cert. No.: 00026
Sequential File No.:	00481	Lin. Date: 01/04/2017
		Lin. No.: 00026
		Soln. Date: 05/25/2017
		Soln. No.: 00188
		File Date: 05/25/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0080
Solution Control Lot:	16150	Expires: 07/06/2018
		Bottle No.: 1161

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:57D		
Control 1 EC	0.099%	07:57D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:58D		
Control 2 EC	0.099%	07:59D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	07:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:59D		
Control 3 EC	0.099%	08:00D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: P First Name: JESSE MI: R

Signature:  Badge No.: 421
Date: 05/25/2017