

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00524	Calib. Date:	03/27/2018
Certification File No.:	00525	Cert. Date:	03/27/2018
Linearity File No.:	00526	Lin. Date:	03/27/2018
Solution File No.:	00529	Soln. Date:	05/29/2018
Sequential File No.:	00529	File Date:	05/29/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUH S3-0079
Solution Control Lot:	17420	Expires:	12/11/2019
		Bottle No.:	1024

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:36D		
Control 1 EC	0.101%	07:36D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:37D		
Control 2 EC	0.100%	07:38D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:38D		
Control 3 EC	0.099%	07:39D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:40D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Badge No.: 421

Date: 05/29/2018