

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00482 Calib. Date: 05/31/2017 Calib. No.: 00036
Certification File No.: 00483 Cert. Date: 05/31/2017 Cert. No.: 00027
Linearity File No.: 00484 Lin. Date: 05/31/2017 Lin. No.: 00027
Solution File No.: 00491 Soln. Date: 07/26/2017 Soln. No.: 00191
Sequential File No.: 00491 File Date: 07/26/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0079
Control Solution %: 0.100% Expires: 03/20/2019
Solution Control Lot: 17110 Bottle No.: 0826

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:12D		
Control 1 EC	0.104%	12:13D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	12:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:13D		
Control 2 EC	0.103%	12:14D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:14D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:14D		
Control 3 EC	0.102%	12:15D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:15D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 07/26/2017