

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00524	Calib. Date: 03/27/2018
Certification File No.:	00525	Calib. No.: 00038
Linearity File No.:	00526	Cert. Date: 03/27/2018
Solution File No.:	00531	Cert. No.: 00029
Sequential File No.:	00531	Lin. Date: 03/27/2018
		Lin. No.: 00029
		Soln. Date: 07/25/2018
		Soln. No.: 00205
		File Date: 07/25/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0079
Solution Control Lot:	17420	Expires: 12/11/2019
		Bottle No.: 0986

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:20D		
Control 1 EC	0.100%	12:21D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:22D		
Control 2 EC	0.099%	12:22D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:23D		
Control 3 EC	0.099%	12:23D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:24D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_

Badge No.: 421

Date: 07/25/2018