

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00482	Calib. Date:	05/31/2017
Certification File No.:	00483	Calib. No.:	00036
Linearity File No.:	00484	Cert. Date:	05/31/2017
Solution File No.:	00492	Lin. Date:	05/31/2017
Sequential File No.:	00492	Lin. No.:	00027
		Soln. Date:	08/24/2017
		Soln. No.:	00192
		File Date:	08/24/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUH S3-0079
Solution Control Lot:	17110	Expires:	03/20/2019
		Bottle No.:	1053

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:21D		
Control 1 EC	0.101%	07:22D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	07:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:22D		
Control 2 EC	0.100%	07:23D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	07:23D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:23D		
Control 3 EC	0.101%	07:24D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	07:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Badge No.: 421

Date: 08/24/2017