

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTL-0014  
Location: SOUTH BRUNSWICK POLICE  
Calibration File No.: 00524 Calib. Date: 03/27/2018 Calib. No.: 00038  
Certification File No.: 00525 Cert. Date: 03/27/2018 Cert. No.: 00029  
Linearity File No.: 00526 Lin. Date: 03/27/2018 Lin. No.: 00029  
Solution File No.: 00532 Soln. Date: 08/23/2018 Soln. No.: 00206  
Sequential File No.: 00532 File Date: 08/23/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0079  
Control Solution %: 0.100% Expires: 12/11/2019  
Solution Control Lot: 17420 Bottle No.: 0982

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 07:53D |                |                     |
| Control 1 EC      | 0.099% | 07:53D | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.099% | 07:53D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:54D |                |                     |
| Control 2 EC      | 0.098% | 07:55D | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.099% | 07:55D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:55D |                |                     |
| Control 3 EC      | 0.098% | 07:56D | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.100% | 07:56D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 07:57D |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: \_\_\_\_\_

 #421

Badge No.: 421

Date: 08/23/2018

# Calibrating Unit

## New Standard Solution Report

|                         |                        |              |              |
|-------------------------|------------------------|--------------|--------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C  | Serial No.:  | ARTL-0014    |
| Location:               | SOUTH BRUNSWICK POLICE |              |              |
| Calibration File No.:   | 00524                  | Calib. Date: | 03/27/2018   |
| Certification File No.: | 00525                  | Cert. Date:  | 03/27/2018   |
| Linearity File No.:     | 00526                  | Lin. Date:   | 03/27/2018   |
| Solution File No.:      | 00533                  | Soln. Date:  | 08/23/2018   |
| Sequential File No.:    | 00533                  | File Date:   | 08/23/2018   |
|                         |                        |              |              |
| Calibrating Unit:       | WET                    | Model No.:   | CU-34        |
| Control Solution %:     | 0.100%                 | Serial No.:  | DDUH S3-0079 |
| Solution Control Lot:   | 17420                  | Expires:     | 12/11/2019   |
|                         |                        | Bottle No.:  | 0982         |

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 11:47D |                |                     |
| Control 1 EC      | 0.099% | 11:48D | 34.0°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.100% | 11:48D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:48D |                |                     |
| Control 2 EC      | 0.098% | 11:49D | 34.0°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.098% | 11:49D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:50D |                |                     |
| Control 3 EC      | 0.098% | 11:50D | 34.0°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.099% | 11:50D | 34.0°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 11:51D |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BLAKE

First Name: JESSE

MI: R

Signature:  # 421

Badge No.: 421

Date: 08/23/2018