

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTL-0014
Location: SOUTH BRUNSWICK POLICE
Calibration File No.: 00524 Calib. Date: 03/27/2018 Calib. No.: 00038
Certification File No.: 00525 Cert. Date: 03/27/2018 Cert. No.: 00029
Linearity File No.: 00526 Lin. Date: 03/27/2018 Lin. No.: 00029
Solution File No.: 00534 Soln. Date: 08/24/2018 Soln. No.: 00208
Sequential File No.: 00534 File Date: 08/24/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUH S3-0079
Control Solution %: 0.100% Expires: 12/11/2019
Solution Control Lot: 17420 Bottle No.: 1021

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:55D		
Control 1 EC	0.100%	11:56D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:57D		
Control 2 EC	0.098%	11:57D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:58D		
Control 3 EC	0.098%	11:58D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:59D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

 #421

Badge No.: 421

Date: 08/24/2018