

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00482	Calib. Date: 05/31/2017
Certification File No.:	00483	Calib. No.: 00036
Linearity File No.:	00484	Cert. Date: 05/31/2017
Solution File No.:	00494	Cert. No.: 00027
Sequential File No.:	00494	Lin. Date: 05/31/2017
		Lin. No.: 00027
		Soln. Date: 09/22/2017
		Soln. No.: 00193
		File Date: 09/22/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0079
Solution Control Lot:	17110	Expires: 03/20/2019
		Bottle No.: 1011

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:37D		
Control 1 EC	0.101%	13:38D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:38D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:38D		
Control 2 EC	0.100%	13:39D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:39D		
Control 3 EC	0.100%	13:40D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Badge No.: 421

Date: 09/22/2017