

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0014
Location:	SOUTH BRUNSWICK POLICE		
Calibration File No.:	00537	Calib. Date:	09/05/2018
Certification File No.:	00538	Cert. Date:	09/05/2018
Linearity File No.:	00539	Lin. Date:	09/05/2018
Solution File No.:	00541	Soln. Date:	10/04/2018
Sequential File No.:	00541	File Date:	10/04/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUH S3-0080
Solution Control Lot:	17420	Expires:	12/11/2019
		Bottle No.:	0993

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:12D		
Control 1 EC	0.101%	13:12D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:13D		
Control 2 EC	0.100%	13:13D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	13:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:14D		
Control 3 EC	0.101%	13:15D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	13:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Badge No.: 421

Date: 10/04/2018