

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00537	Calib. Date: 09/05/2018
Certification File No.:	00538	Calib. No.: 00039
Linearity File No.:	00539	Cert. Date: 09/05/2018
Solution File No.:	00542	Cert. No.: 00030
Sequential File No.:	00542	Lin. Date: 09/05/2018
		Lin. No.: 00030
		Soln. Date: 11/08/2018
		Soln. No.: 00211
		File Date: 11/08/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0080
Solution Control Lot:	18150	Expires: 05/10/2020
		Bottle No.: 1353

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:17S		
Control 1 EC	0.100%	08:18S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	08:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:18S		
Control 2 EC	0.100%	08:19S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:20S		
Control 3 EC	0.100%	08:20S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:21S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 11/08/2018