

Calibrating Unit

New Standard Solution Report

| | | |
|-------------------------|------------------------|--------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARTL-0014 |
| Location: | SOUTH BRUNSWICK POLICE | |
| Calibration File No.: | 00497 | Calib. Date: 11/24/2017 |
| Certification File No.: | 00498 | Cert. No.: 00037 |
| Linearity File No.: | 00499 | Cert. Date: 11/24/2017 |
| Solution File No.: | 00501 | Lin. No.: 00028 |
| Sequential File No.: | 00501 | Soln. No.: 00197 |
| | | File Date: 12/20/2017 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDUH S3-0080 |
| Solution Control Lot: | 17110 | Expires: 03/20/2019 |
| | | Bottle No.: 0911 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 08:11S | | |
| Control 1 EC | 0.102% | 08:11S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.101% | 08:11S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:12S | | |
| Control 2 EC | 0.100% | 08:13S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 08:13S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:13S | | |
| Control 3 EC | 0.100% | 08:14S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 08:14S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:15S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____

Badge No.: 421

Date: 12/20/2017