

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0014
Location:	SOUTH BRUNSWICK POLICE	
Calibration File No.:	00584	Calib. No.: 00043
Certification File No.:	00585	Cert. No.: 00034
Linearity File No.:	00586	Lin. No.: 00034
Solution File No.:	00592	Soln. No.: 00243
Sequential File No.:	00592	File Date: 10/09/2020
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUH S3-0080
Solution Control Lot:	19020	Expires: 01/08/2021
		Bottle No.: 0285

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:59D		
Control 1 EC	0.098%	11:00D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:01D		
Control 2 EC	0.099%	11:01D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	11:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:02D		
Control 3 EC	0.100%	11:03D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	11:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BLAKE

First Name: JESSE

MI: R

Signature: _____



Badge No.: 421

Date: 10/09/2020